

A SELECTION FROM THE VAST RANGE OF

ClarkeTM

QUALITY PRODUCTS

AIR COMPRESSORS
From DIY to industrial. Plus air tools, spray guns and accessories.

GENERATORS
Prime duty or emergency standby for business, home and leisure.

POWER WASHERS
Hot and cold, electric and engine driven - we have what you need.

WELDERS
Mig, Arc, Tig and Spot. From DIY to automotive.

METALWORKING
Drills, grinders and saws for DIY and professional use.

WOODWORKING
Saws, sanders, lathes, mortisers and dust extraction.

HYDRAULICS
Cranes, body repair kits, transmissions for all types of workshop use.

WATER PUMPS
Submersible, electric and engine driven for DIY, agriculture and industry.

POWER TOOLS
Angle grinders, cordless drill sets, saws and sanders.

STARTER/CHARGERS
All sizes for car & commercial use.

Clarke INTERNATIONAL
For spare parts and servicing, please contact your nearest dealer, or Clarke International on
020 - 8988 - 7400
e-mail: Parts@clarkeinternational.com e-mail: Service@clarkeinternational.com

Clarke[®]

DIGITAL LASER TACHOMETER



MODEL No.
CT1

OPERATING INSTRUCTIONS

0507



When disposing of this product, do not dispose of with general waste. It must be disposed of according to the laws governing Waste Electrical and Electronic equipment, at a recognised disposal facility.

Clarke®

2



Clarke®

7

BATTERY REPLACEMENT

When the batteries need replacement, the symbol 'LO' will appear on the digital display.

To replace the batteries, slide back the cover on the rear of the unit and remove the spent batteries.

Pay particular attention to polarity when replacing the batteries as incorrect polarity could cause permanent, serious damage to the instrument.

If the instrument is not to be used for some time, it is strongly advised to remove the batteries.

PARTS & SERVICE CONTACTS

For Spare Parts and Service, please contact your nearest dealer, or CLARKE International, on one of the following numbers.

PARTS & SERVICE TEL: 020 8988 7400

PARTS & SERVICE FAX: 020 8558 3622

or e-mail as follows:

PARTS: Parts@clarkeinternational.com

SERVICE: Service@clarkeinternational.com

Thank you for purchasing this CLARKE Digital Laser Tachometer.

Please read this booklet thoroughly. Your Clarke Tachometer will give excellent service if it is used carefully and in accordance with the following advice and recommendations. Never use excessive force and as with all tools, it should be treated with care and respect.

GUARANTEE

This CLARKE product is guaranteed against faulty manufacture for a period of 12 months from the date of purchase. Please keep your receipt as proof of purchase. This guarantee is invalid if the product is found to have been abused or tampered with in any way, or not used for the purpose for which it was intended.

Faulty goods should be returned to their place of purchase, no product can be returned to us without prior permission. This guarantee does not effect your statutory rights.

SPECIFICATIONS

Model No.	CT1
Part No.	4501110
Measuring Range	5 - 99,999 RPM
Resolution	0.1RPM up to 1000RPM 1 RPM above 1000RPM
Accuracy	±(0.05% + 1 digit)
Laser Class	II
Operating Temperature	0 - 50°C (32 - 132°F)
Dimensions	190x72x32mm
Sampling Time	1 sec (over 6RPM)
Weight	0.2kg
Voltage	6V
Batteries	4xAA

FEATURES

Your Digital Laser Tachometer provides a wide measuring range at high resolution, using a large LCD display.

An exclusive microchip, LSI-circuit and crystal time base provide a highly accurate measurement with a fast measuring time.

The last value Max and Min values are stored in memory and may be recalled by pressing a memory button.

Self adhesive reflectors are provided for the surface of the object to be measured.

The unit is constructed from light, strong, durable components, ensuring long and maintenance free service, and is designed to fit comfortably into the palm of the hand.



A - Operation Button

C - Memory Recall Button

B - LCD Display

D - Target Indicator

Clarke®

MEASURING PROCEDURE

1. RPM Measurement

- 1.1 Apply the reflecting mark to the surface of the object to be measured.
- 1.2 With the object rotating, depress the Operation Button and align the visible light beam with the reflecting mark, as it passes the beam.

Check to ensure the 'Target Indicator', on the display, illuminates when the reflector passes through the light beam.

- 1.3 When the reading stabilises, release the Operation Button.

NOTE:

If the revolutions being measured are very low, i.e. less than 50RPM, it is advisable to stick more reflectors at regular (measured) intervals on the surface of the object.

To obtain a correct reading, simply divide the display reading by the numbers of reflectors.

2. Memory Recall Button

The last measurement, Maximum and Minimum values are stored in memory, that is, the last values before the Operation Button was released. To recall the values, proceed as follows:

- 3.1 Press the Memory button once. The symbols 'LA' will be displayed along with the LAST VALUE
- 3.2 Press the Memory Button a second time and the symbols 'UP' will be displayed together with the MAXIMUM VALUE of the last measurement.
- 3.3 Press the Memory Button a third time and the symbols 'dn' will be displayed together with the MINIMUM VALUE of the last measurement.

Clarke®